Estimation of Affine Transformations directly from Tomographic Projections

René Mooser, Erwin Hack, Urs Sennhauser
 Swiss Federal Laboratories for Materials Testing and Research (EMPA)
 Electronics/Metrology Laboratory
 Ueberlandstrasse 129, CH-8600 Duebendorf
 {rene.mooser,erwin.hack,urs.sennhauser}@empa.ch

Gábor Székely
 Swiss Federal Institute of Technology Zurich
 Computer Vision Laboratory
 Sternwartstrasse 7, CH-8092 Zurich
 szekely@vision.ee.ethz.ch

Abstract

This paper presents a new approach to estimate affine transformations from tomographic projections. Instead of estimating the deformation from the reconstructed data, we introduce a method which works directly in the projection data. Using a parallel beam geometry, we show that any affine transformation results in scaling and translation of the projection data. Furthermore, we present a method which recovers the parameters of the deformation. Numerical experiments conclude the paper.

1. Introduction

X-ray Computed Tomography (CT) is a tool widely recognized and used for various applications in materials science and medicine. It is often of great interest to examine the behavior of a specific sample under load, e. g. the deformation induced by applying stress to the sample. The estimated deformation can then be used to study mechanical properties like strain [7], which is important e.g. in clinical applications [16]. Digital motion estimation is a key tool to find and determine deformations from digital images as produced by CT.

Here, we propose a method to determine the affine deformation of a sample directly from the projections, thus avoiding the computationally expensive reconstruction which would otherwise be needed. Approaches for estimating deformations directly in the projections have recently received a lot of attention [2, 6, 11, 14]. However, existing work is limited to rigid-body transforms (i.e. rotations and translations) and uniform scaling. Our method incorporates all degrees of freedom of an affine transformation, that is, rotation, translation, non-uniform scaling and non-uniform shearing. For the purpose of this paper, we restrict ourselves to two-

dimensional parallel geometry and absorption contrast X-ray CT.

2. X-Ray Computed Tomography and the Radon transform

Consider an intensity image given by its values $f(\mathbf{x})$ defined for $\mathbf{x} \in \Omega \subset \mathbb{R}^2$. The Radon transform of $f(\mathbf{x})$ is defined [4] as the mapping

$$\begin{aligned}\check{f}(p, \xi_\theta) &= \mathcal{R}_\theta[f(\mathbf{x}), p] \\ &= \iint_{\Omega} f(\mathbf{x}) \delta(p - \langle \xi_\theta, \mathbf{x} \rangle) d\mathbf{x}\end{aligned}\quad (1)$$

where $\xi_\theta = \begin{pmatrix} \cos(\theta) \\ \sin(\theta) \end{pmatrix}$ is a unit vector, δ is the Dirac delta function and $\langle \cdot, \cdot \rangle$ denotes the standard inner product on $\mathbb{R}^2$. Let

$$\Theta = \{\theta_1, \theta_2, \dots, \theta_P\} \quad (2)$$

be a set of projection angles, $0 \leq \theta_i < 180^\circ$. For a given set of coordinates p and a given set of angles Θ , we call the set of all projections obtained by (1) a Sinogram. Fig. 1 shows a Sinogram of a simple image, Fig. 2 shows the geometry setup.

2.1. X-ray Computed Tomography

In X-ray CT, the investigated sample is placed between an X-ray source and a detector device, which measures the intensity of the transmitted X-rays I_{tr} . By rotating the object in small steps $\Delta\theta$, an absorption contrast image is generated for every angle θ . For a monochromatic beam with incident intensity I_0 , the normalized measured intensity $\frac{I_{tr}}{I_0}$ is

$$\frac{I_{tr}}{I_0} = \exp\left(-\int_L \mu(s) ds\right) \quad (3)$$

Figure 1. Radon transform of an image. The (x, y) coordinate system of the image (a) is mapped to the (θ, p) coordinate system (b).

Figure 2. Parallel beam geometry setup for a given rotation angle θ .

where L is a projection line through the object and $\mu(s)$ is the material dependent attenuation coefficient. Using the projection geometry given in Fig. 2, a projection line L is defined by

$$p = \langle \mathbf{x}, \xi_\theta \rangle \quad (4)$$

and we see that equation (3) formally equals equation (1) by taking the log on both sides of (3). Therefore, in digital images, the image values $f(\mathbf{x})$ can be related to the attenuation coefficient $\mu(s)$. This shows that a discrete implementation of equation (1) can be used to simulate X-ray CT [9].

3. Affine transformations and the Radon transform

As already mentioned, affine transformations consist of rotation, translation, scaling and shearing. Using matrices,

we can therefore construct affine transformations with

$$\mathbf{T}_a(\mathbf{x}) = \mathbf{A}_a \mathbf{x} + \mathbf{t} \quad (5)$$

where

$$\mathbf{A}_a = \mathbf{A}_{rot} \cdot \mathbf{A}_{scale} \cdot \mathbf{A}_{shear} \quad (6)$$

and

$$\mathbf{t} = \begin{pmatrix} t_x \\ t_y \end{pmatrix} \quad (7)$$

$$\mathbf{A}_{rot} = \begin{pmatrix} \cos(\varphi) & -\sin(\varphi) \\ \sin(\varphi) & \cos(\varphi) \end{pmatrix} \quad (8)$$

$$\mathbf{A}_{scale} = \begin{pmatrix} sx & 0 \\ 0 & sy \end{pmatrix} \quad (9)$$

$$\mathbf{A}_{shear} = \begin{pmatrix} 1 & shx \\ shy & 1 \end{pmatrix} \quad (10)$$

The matrix multiplications in (6) are generally not pairwise commutative, such that the order of multiplication will affect the resulting transformation.

3.1. Radon transform under affine transformation

We now consider the Radon transform of an affine deformed object f , that is,

$$\begin{aligned} \mathcal{R}_\theta[f(\mathbf{T}_a(\mathbf{x})), p] &= \mathcal{R}_\theta[f(\mathbf{A}_a \mathbf{x} + \mathbf{t}), p] \\ &= \iint_{\Omega} f(\mathbf{A}_a \mathbf{x} + \mathbf{t}) \delta(p - \langle \xi_\theta, \mathbf{x} \rangle) d\mathbf{x} \end{aligned} \quad (11)$$

Setting $\mathbf{y} = \mathbf{A}_a \mathbf{x} + \mathbf{t}$ and assuming that $\mathbf{A}_a$ is non-singular, we get

$$\begin{aligned} \mathcal{R}_\theta[f(\mathbf{T}_a(\mathbf{x})), p] &= |\det(\mathbf{B}_a)| \iint_{\Omega} f(\mathbf{y}) \delta\left(p - \langle \mathbf{B}_a^\top \xi_\theta, \mathbf{y} - \mathbf{t} \rangle\right) d\mathbf{y} \\ &= |\det(\mathbf{B}_a)| \tilde{f}\left(p + \langle \mathbf{B}_a^\top \xi_\theta, \mathbf{t} \rangle, \mathbf{B}_a^\top \xi_\theta\right) \end{aligned} \quad (12)$$

where we have used $\mathbf{B}_a = \mathbf{A}_a^{-1}$.

As $\mathbf{B}_a^\top \xi_\theta$ is not generally a unit vector anymore, we have to divide (12) by $\|\mathbf{B}_a^\top \xi_\theta\|$, resulting in

$$\begin{aligned} \mathcal{R}_\theta[f(\mathbf{T}_a(\mathbf{x})), p] &= \frac{|\det(\mathbf{B}_a)|}{\|\mathbf{B}_a^\top \xi_\theta\|} \tilde{f}\left(\frac{p + \langle \mathbf{B}_a^\top \xi_\theta, \mathbf{t} \rangle}{\|\mathbf{B}_a^\top \xi_\theta\|}, \frac{\mathbf{B}_a^\top \xi_\theta}{\|\mathbf{B}_a^\top \xi_\theta\|}\right) \\ &= \tilde{f}(\tilde{\mathbf{T}}_a(p), \tilde{\mathbf{T}}_a(\xi_\theta)), \end{aligned} \quad (13)$$

where we have introduced the notation $\tilde{f}(\tilde{\mathbf{T}}_a(p), \tilde{\mathbf{T}}_a(\xi_\theta))$ to express the action of the affine mapping $\mathbf{T}_a$ on the Radon transform $\tilde{f}(p, \xi_\theta)$. Equation (13) has a very interesting and compelling interpretation: It basically means that the Radon transform (and hence a tomographic scan) of an affine deformed object f can be written as the tomographic scan of the undeformed object, using a deformed scanning geometry. As it is well known, affine transformations map parallel lines to parallel lines, which means that the deformed scanning geometry will again be a parallel beam geometry [5].

3.2. Detailed Analysis

We now give a detailed analysis of the effects of $\mathbf{B}_a$ and $\mathbf{t}$ on the deformed projections $\check{f}(\check{\mathbf{T}}_a(p), \check{\mathbf{T}}_a(\xi_\theta))$. First, we notice that the mapping $p \mapsto p + \langle \mathbf{B}_a^\top \xi_\theta, \mathbf{t} \rangle$ is simply a translation with respect to p . The division by $\|\mathbf{B}_a^\top \xi_\theta\|$ can be interpreted as a scaling of $p + \langle \mathbf{B}_a^\top \xi_\theta, \mathbf{t} \rangle$. We therefore conclude that an affine transformation will generally affect the first variable of $\check{f}(\check{\mathbf{T}}_a(p), \check{\mathbf{T}}_a(\xi_\theta))$ by scaling and translating it. Doing the same analysis for the second variable of $\check{f}(\check{\mathbf{T}}_a(p), \check{\mathbf{T}}_a(\xi_\theta))$ reveals that the mapping $\xi_\theta \mapsto \frac{\mathbf{B}_a^\top \xi_\theta}{\|\mathbf{B}_a^\top \xi_\theta\|}$ is a rotation of the vector ξ_θ around the origin of the coordinate system (see Fig. 3), the rotation angle is determined via

$$\gamma = \arccos \left(\frac{\langle \xi_\theta, \mathbf{B}_a^\top \xi_\theta \rangle}{\|\mathbf{B}_a^\top \xi_\theta\|} \right). \quad (14)$$

As it can be seen from Fig. 1, a rotation of the scanning

Figure 3. Deformation of the scanning geometry due to affine transformation. The scaling of p and the rotation of ξ_θ (rotation angle γ) is shown. Affine parameters are $sx = sy = 1.8$, $tx = ty = 0$, $shx = shy = 1$, $\varphi = 40^\circ$.

geometry is simply a translation with respect to θ in the sinogram.

This implies that we need only three basic operations in the sinogram domain (translation and scaling of p , translation of θ) to describe the effect of an arbitrary affine transformation in the image domain. Until now, we have neglected the influence of the constant factor $\frac{|\det(\mathbf{B}_a)|}{\|\mathbf{B}_a^\top \xi_\theta\|}$ in (13), we will elaborate in the next section why we do not need to take this factor into account for our further arguments.

4. Deformation estimation procedure

For the following arguments, we assume that we have two samples (reference and deformed), given by their (unknown) density distributions $f_{ref}(\mathbf{x})$ and $f_{def}(\mathbf{x})$, with $f_{def}(\mathbf{x}) = f_{ref}(\mathbf{T}_a(\mathbf{x}))$. We acquire tomographic scans of f_{ref} and f_{def} , resulting in the Sinograms S_{ref} and S_{def} . Our goal is now to estimate the parameters of $\mathbf{T}_a$, using only the information given in S_{ref} and S_{def} .

4.1. Iterative estimation

Let Θ_{ref} and Θ_{def} be the projection angles used to acquire S_{ref} and S_{def} , respectively. We want to minimize a target function given by

$$\Psi(\mathbf{T}_a) = \frac{1}{2} \sum_{\theta \in \Theta_{ref}} (\psi_\theta(\mathbf{T}_a))^2, \quad (15)$$

where

$$\psi_\theta(\mathbf{T}_a) = \mathcal{M}(\check{f}_{ref}(\check{\mathbf{T}}_a(p), \check{\mathbf{T}}_a(\xi_\theta)), \check{f}_{def}(p, \xi_\theta)) \quad (16)$$

and $\mathcal{M}(I_1, I_2)$ is a similarity measure which is minimal if $I_1 = I_2$.

We choose to use normalized cross correlation [8, 10] as similarity measure, defined as

$$\mathcal{M}(I_1, I_2) = 1 - \frac{\sum_{\mathbf{x}} (I_1(\mathbf{x}) - \bar{I}_1)(I_2(\mathbf{x}) - \bar{I}_2)}{\left(\sum_{\mathbf{x}} (I_1(\mathbf{x}) - \bar{I}_1)^2 \sum_{\mathbf{x}} (I_2(\mathbf{x}) - \bar{I}_2)^2 \right)^{\frac{1}{2}}}. \quad (17)$$

For all I_1, I_2 , $\mathcal{M}(I_1, I_2) \in [0 \dots 2]$.

Normalized cross correlation has the advantage that it is invariant under linear transformations, i.e. for $I_2 = mI_1 + b$ it holds that $\mathcal{M}(I_1, I_2) = 0$ for all $m \in \mathbf{R} \setminus \{0\}, b \in \mathbf{R}$. Therefore, this allows us to ignore the constant factor $\frac{|\det(\mathbf{B}_a)|}{\|\mathbf{B}_a^\top \xi_\theta\|}$ in (13), which simplifies the optimization process.

The estimation procedure then reads

$$\text{find } \mathbf{T}_a : \Psi(\mathbf{T}_a) \rightarrow \min. \quad (18)$$

This is the description of a non-linear least squares problem, which may be solved e.g. by the well known Gauss-Newton method [1]. The Gauss-Newton method starts with an initial estimate $\mathbf{T}_a^0$ and then iteratively creates a sequence $\mathbf{T}_a^1, \mathbf{T}_a^2, \dots, \mathbf{T}_a^N$ such that $\Psi(\mathbf{T}_a^{j+1}) < \Psi(\mathbf{T}_a^j)$. Typically, we choose $\mathbf{T}_a^0$ to be the identity, i.e. $\mathbf{T}_a^0 = \begin{pmatrix} 1 & 0 \\ 0 & 1 \end{pmatrix} + \begin{pmatrix} 0 \\ 0 \end{pmatrix}$. We abort the estimation procedure if the difference between two estimates is below a given threshold tol .

4.2. Smoothed cubic spline interpolation

In order to evaluate expression (16), we must know $\check{f}_{ref}(\check{\mathbf{T}}_a(p), \check{\mathbf{T}}_a(\xi_\theta))$ for every value of $\check{\mathbf{T}}_a(p)$ and $\check{\mathbf{T}}_a(\xi_\theta)$.

As the number of projections angles in Θ_{ref} and the resolution of the detector coordinate p are generally limited, we use a cubic smoothing spline [3] to interpolate values which are not directly accessible. The smoothing ensures that high frequency oscillations due to noise (e.g. detector noise) are damped and their influence on the optimization procedure is limited.

5. Experimental procedure

We will now describe the experimental setup and give a description of the error measures we use to evaluate our method.

5.1. Simulation of the tomographic projections

We use expression (3) as a starting point for our simulation of the tomographic projections. We substitute the integral in (3) by a discrete sum and account for typical CCD photon counting noise by replacing the deterministic values of I_{tr} by Poisson distributed values, i.e.

$$\tilde{I}_{tr} \sim \text{Pois} \left(I_0 \exp(-k \int_L f(s) ds) \right). \quad (19)$$

The constant $k \in \mathbf{R}$ is chosen such that the minimum transmission (i.e. the minimum value of I_{tr}) equals a given percentage of the maximum transmission. Here, we choose the minimum transmission to be 20% of the maximum transmission. We may now use I_0 (the number of incident photons per detector pixel), to simulate any desired signal to noise ratio (SNR), which is defined as [13]

$$\text{SNR}(dB) = 10 \log_{10} \left(\frac{\sigma_c^2}{\sigma_n^2} \right), \quad (20)$$

where σ_c^2 and σ_n^2 are the variances of a noise free projection and the noise, respectively. Finally, we rescale $\tilde{I}_{tr}$, using

$$\tilde{I}_{tr}^{scale} = -\frac{1}{k} \ln \left(\frac{\tilde{I}_{tr}}{I_0} \right). \quad (21)$$

For a given image $f(\mathbf{x})$ and a given set of projection angles Θ , the simulated Sinogram therefore consists of $\tilde{I}_{tr}^{scale}$, computed at the angles $\theta \in \Theta$. We illustrate the effect of the choice of I_0 in Fig. 4, where we show reconstructions for different I_0 . For these images, the number of projection angles is 180, evenly spaced between $[0^\circ, 180^\circ]$, the reconstruction routine is a standard filtered backprojection [12].

5.2. Test image

We selected the famous Shepp-Logan head phantom, defined in a modified version in [15] as test image. All pixel values are in the range $[0, 1]$, and higher values stand for higher absorption coefficients. As image size, we choose $N_x \times N_y = 512 \times 512$ pixels, but our proposed method does not depend on the image size. Fig. 4(a) shows the test image.

Figure 4. Effect of the parameter I_0 on the image quality. Noise free reconstruction (a), $I_0 = 1 \cdot 10^3$, $\text{SNR} \approx 21\text{dB}$ (b), $I_0 = 1 \cdot 10^2$, $\text{SNR} \approx 11\text{dB}$ (c).

5.3. Error measures

Let f_{ref} be the undeformed image, $\mathbf{T}_a$ the transformation and $f_{def} = f_{ref}(\mathbf{T}_a(\mathbf{x}))$ the deformed image. We then compute an estimated transformation $\mathbf{T}_a^{est}$ using the procedure described in section 4. Once we have an estimate $\mathbf{T}_a^{est}$ for the transformation $\mathbf{T}_a$, we are able to calculate an estimated deformation vector field $\mathbf{v}_{est}$ at every pixel of the test image and compare it to the reference deformation field $\mathbf{v}_{ref}$. We choose the origin of the coordinate system to be in the center of the image and define the mean angular error as

$$\Lambda_{ang} = \frac{1}{N_x N_y} \sum_{\mathbf{x}} \arccos \left(\frac{\langle \mathbf{v}_{ref}(\mathbf{x}), \mathbf{v}_{est}(\mathbf{x}) \rangle}{\|\mathbf{v}_{ref}(\mathbf{x})\| \|\mathbf{v}_{est}(\mathbf{x})\|} \right) \quad (22)$$

Additionally, we also use normalized cross correlation (17) as an error measure. For this, we calculate the normalized cross correlation coefficient between the test image deformed with $\mathbf{T}_a^{est}$ and the test image deformed with $\mathbf{T}_a$, i.e.

$$\Lambda_{ncc} = \mathcal{M}(\mathbf{T}_a^{est}(f_{ref}), \mathbf{T}_a(f_{ref})). \quad (23)$$

We also compute

$$\Lambda_{ncc,nodef} = \mathcal{M}(f_{ref}, \mathbf{T}_a(f_{ref})) \quad (24)$$

which enables us to quantify the improvement in the cross correlation error (23) achieved by the transformation estimation.

6. Results

In order to get statistically relevant results, we use the parameter I_0 in (19) as a steering parameter to produce a desired SNR. Here, we choose $I_0 = \{10^2, 10^3, 10^4, 10^5\}$. For every SNR, we create 100 varying affine deformations by randomly selecting the parameters $t_x, t_y, s_x, s_y, sh_x, sh_y, \varphi$ and evaluate the error measures (22), (23) and (24). The ranges from which we select the affine parameters are given in table 1, each value has an uniform probability to be selected. Also, the number of projection angles used for the simulation of the tomographic projections is 90 to compute S_{ref} and 180 to

Table 1. Valid ranges for the affine parameters

Parameter:	Range:
tx, ty (pixel)	$[-5, 5]$
sx, sy	$[0.9, 1.1]$
shx, shy	$[-0.1, 0.1]$
φ (deg)	$[-10^\circ, 10^\circ]$

Figure 5. Λ_{ncc} compared to $\Lambda_{nnc,nodef}$ for different SNR (a). Mean angular error Λ_{ang} for different SNR (b). In both cases, the standard deviation is also plotted.

compute S_{def} , all evenly spaced between $[0^\circ, 180^\circ]$. The estimation results are plotted in Fig. 5. First, we notice that the SNR has only a weak influence on the estimation accuracy, which is most likely due to the use of smoothing splines and the robustness of the normalized cross correlation. Also, we see that the standard deviation of the angular error is quite large, which indicates the performance of our method does depend on the type of deformation applied to the sample. On the other hand, the improvement of the cross correlation value is clearly visible, the achieved value is (except for the lowest SNR) always well below 0.2. Fig.

6 shows an example of the estimation results. We did not

Figure 6. Example of the estimation results. Reconstructed test image deformed with the ground truth deformation (a). Reconstructed test image deformed with the estimated deformation (b). Pointwise absolute difference, scaled by a factor of 3 for better visibility (c). The following parameters have been used for this example: $tx = 4, ty = -3, sx = 1.15, sy = 1, shx = 0.1, shy = 0.1, \varphi = 0^\circ, SNR = 21dB$.

perform any extended evaluation concerning the cpu time of the estimation procedure, but the Gauss-Newton procedure roughly needs 15 iterations to reach the convergence tolerance tol and this takes about 2 minutes on a modern workstation. However, there is much potential for further optimization, and we guess that a more efficient implementation may need less than 10 seconds to converge.

7. Conclusion

We have shown that it is possible to estimate global affine deformations directly from tomographic projections, avoiding the computationally expensive reconstruction. We evaluated our results by comparison to known ground truth data, and observed that our method produces accurate estimates even at low SNR. In our simulations, we have accounted for photon counting noise and included the possibility to adjust the number of incoming photons, thus enabling us for reasonably accurate simulations. For further research, it would be interesting to study other scanning geometries like fan-beam or cone-beam. Also, the question whether the estimation accuracy is better in the projection domain or in the image domain (using reconstructed images) is a valuable topic for further efforts.

References

- [1] A. Björck, *Numerical methods for least squares problems*, Society for Industrial and Applied Mathematics, Philadelphia, 1996.
- [2] H. F. Chen, D. Z. Yao, R. H. Li, and W. F. Chen, "A projection-based image registration algorithm and its application", *Brain Topography*, 18(1), 2005, pp. 47–58.
- [3] C. De Boor, *A practical guide to splines*, Springer, New York, rev. edition, 2001.
- [4] S. R. Deans, *The Radon transform and some of its applications*, Wiley, New York a.o., 1983.
- [5] L. Desbat, S. Roux, and P. Grangeat, "Compensation of some time dependent deformations in tomography", *IEEE Transactions on Medical Imaging*, 26(2), 2007, pp. 261–269.
- [6] E. E. Fitchard, J. S. Aldridge, K. Ruchala, G. Fang, J. Balog, D. W. Pearson, G. H. Olivera, E. A. Schloesser, D. Wenman, P. J. Reckwerdt, and T. R. Mackie, "Registration using tomographic projection files", *Physics in Medicine and Biology*, 44(2), 1999, pp. 495–507.
- [7] M. G. D. Geers, R. DeBorst, and W. A. M. Brekelmans, "Computing strain fields from discrete displacement fields in 2d-solids", *International Journal of Solids and Structures*, 33(29), 1996, pp. 4293–4307.
- [8] A. Giachetti, "Matching techniques to compute image motion", *Image and Vision Computing*, 18(3), 2000, pp. 247–260.
- [9] A. C. Kak and M. Slaney, *Principles of computerized tomographic imaging*, IEEE Press The Institute of Electrical and Electronics Engineers, New York, 1988.
- [10] J. P. Lewis, "Fast normalized cross-correlation", In *Vision Interface*, pp. 120–123. Canadian Image Processing and Pattern Recognition Society, 1995.
- [11] W. G. Lu and T. R. Mackie, "Tomographic motion detection and correction directly in sinogram space", *Physics in Medicine and Biology*, 47(8), 2002, pp. 1267–1284.
- [12] F. Natterer and F. Wuebbeling, *Mathematical methods in image reconstruction*, Society for Industrial and Applied Mathematics, Philadelphia, Pa., 2001.
- [13] D. Robinson and P. Milanfar, "Fast local and global projection-based methods for affine motion estimation", *Journal of Mathematical Imaging and Vision*, 18(1), 2003, pp. 35–54.
- [14] S. Sarkar, M. A. Oghabian, I. Mohammadi, A. Mohammadpour, and A. Rahmim, "A linogram/sinogram cross-correlation method for motion correction in planar and spect imaging", *IEEE Transactions on Nuclear Science*, 54(1), 2007, pp. 71–79.
- [15] P. Toft, *The Radon transform theory and implementation*, Department of Mathematical Modelling, Section for Digital Signal Processing, [Lyngby], 1996.
- [16] E. Verhulp, B. van Rietbergen, and R. Huiskes, "A three-dimensional digital image correlation technique for strain measurements in microstructures", *Journal of Biomechanics*, 37(9), 2004, pp. 1313–1320.

ISPA 2009

Proceedings of the 6th International Symposium on
Image and Signal Processing and Analysis
Salzburg, Austria, September 16-18, 2009

Peter Zinterhof, Sven Lončarić, Andreas Uhl, Alberto Carini (Eds.)

Co-Sponsored by:

EURASIP - European Association for Signal Processing

IEEE Region 8

IEEE Austria Section

Co-Organised by:

University of Salzburg, Austria

Faculty of Electrical Engineering and Computing, University of
Zagreb, Croatia

Volume Editors

Peter Zinterhof

University of Salzburg,
Kapitelgasse 4-6,
AT-5020 Salzburg, Austria
E-mail: peter.zinterhof@sbg.ac.at

Andreas Uhl

University of Salzburg,
Kapitelgasse 4-6,
AT-5020 Salzburg, Austria
E-mail: uhl@cosy.sbg.ac.at

Sven Lončarić

Faculty of Electrical Engineering and Computing,
University of Zagreb
Unska 3, HR-10000 Zagreb, Croatia
E-mail: sven.loncaric@fer.hr

Alberto Carini

Information Science and Technology Institute
University of Urbino "Carlo Bo",
61029 Urbino (PU), Italy
Email: alberto.carini@uniurb.it

ISBN 978-953-184-134-4

IEEE Catalog Number CFP09504-CDR

Additional copies may be ordered from:

IEEE Publications
www.ieee.org

Published by:

University of Salzburg,
Kapitelgasse 4-6, 5020
Salzburg, Austria

and

Faculty of Electrical Engineering and Computing,
University of Zagreb, Unska 3, 10000 Zagreb,
Croatia

This work is subject to copyright. All rights are reserved, whether the whole or part of the material is concerned, specifically the rights of translation, reprinting, re-use of illustrations, recitation, broadcasting, reproduction on microfilms or in any other way, and storage in data banks. No reproduction of any part may take place without the written permission of Faculty of Electrical Engineering and Computing.

© University of Salzburg, Austria and Faculty of Electrical Engineering and Computing, Zagreb, Croatia 2009

Printed at University of Salzburg

Typesetting: Camera-ready by author

ISPA 2009

Editors

P. Zinterhof
S. Lončarić
A. Uhl
A. Carini

CD version

Salzburg, Austria

Proceedings of the
6th International Symposium on

Image and Signal Processing and Analysis

September 16 - 18, 2009
Salzburg, Austria

Organized by

Sponsored by

Sony DADC

Technical support:

secretariat@isispa.org
tel: +385 1 6129911
fax: +385 1 6129652

IEEE CFP09504-CDR
ISBN 953184134-4
ISSN 1845-5956