FIELD CALIBRATION STUDIES FOR IONISATION CHAMBERS IN MIXED HIGH-ENERGY RADIATION FIELDS

C. Theis^{1,*}, D. Forkel-Wirth¹, M. Fuerstner¹, S. Mayer², Th. Otto¹, S. Roesler¹ and H. Vincke¹

¹CERN, CH-1211 Geneva 23, Switzerland

²PSI, CH-5232 Villigen, Switzerland

The monitoring of ambient doses at work places around high-energy accelerators is a challenging task due the complexity of the mixed stray radiation fields encountered. At CERN, mainly Centronics IG5 high-pressure ionisation chambers are used to monitor radiation exposure in mixed fields. The monitors are calibrated in the operational quantity ambient dose equivalent $H^*(10)$ using standard, source-generated photon- and neutron fields. However, the relationship between ionisation chamber reading and ambient dose equivalent in a mixed high-energy radiation field can only be assessed if the spectral response to every component and the field composition is known. Therefore, comprehensive studies were performed at the CERN-EU high-energy reference field facility where the spectral fluence for each particle type has been assessed with Monte Carlo simulations. Moreover, studies have been performed in an accessible controlled radiation area in the vicinity of a beam loss point of CERN's proton synchrotron. The comparison of measurements and calculations has shown reasonable agreement for most exposure conditions. The results indicate that conventionally calibrated ionisation chambers can give satisfactory response in terms of ambient dose equivalent in stray radiation fields at high-energy accelerators in many cases. These studies are one step towards establishing a method of 'field calibration' of radiation protection instruments in which Monte Carlo simulations will be used to establish a correct correlation between the response of specific detectors to a given high-energy radiation field.

INTRODUCTION

Operating high-energy accelerators requires careful monitoring of the radiation environment. Due to the fact that the radiation environment encountered at such accelerators is quite different from the standard uses of such systems it is necessary to perform a thorough investigation of the envisaged devices. Radiation fields around high-energy accelerators like the Large Hadron Collider (LHC) show a composition of many different particle types (neutrons, charged hadrons, leptons and photons) and a wide range of energies from fractions of eV up to several GeV. Consequently, it is necessary to use detectors that respond to various particle types at the same time and exhibit the required sensitivity. At CERN, mainly Centronics IG5 high-pressure ionisation chambers are used to monitor radiation exposure in mixed fields around accelerators. The monitors are calibrated in the operational quantity ambient dose equivalent $H^*(10)$ using standard, source-generated photon- and neutron fields. However, these sources cover only photons in the energy range of 1 MeV (^{137}Cs , ^{60}Co) and neutrons (^{238}Pu –Be) ranging up to 11 MeV, which is much lower than the energies encountered around high-energy accelerators. Therefore, the relationship between detector reading and ambient dose equivalent in a mixed high-energy radiation field can only be assessed if the spectral response of the monitor to every field component is

known. Since the radiation environment around the LHC is not yet available for measurements, studies at existing facilities like the CERN-EU high-energy reference field facility (CERF) or the proton synchrotron (PS) were performed. During these experiments a set of two hydrogen-filled IG5 chambers were used to measure ambient dose equivalent in mixed fields. Accompanying FLUKA Monte Carlo simulations^(1,2) were performed to determine the spectral fluence and the corresponding dose at the exposure locations.

STUDIES AT THE CERF FACILITY

In 1992, the CERF facility was installed in the H6 secondary beam-line of the super PS (SPS) at CERN. A beam with a momentum of 120 GeV/c and a composition of 61% pions, 35% protons and 4% kaons originating from a primary target, was directed at a secondary copper target. This cylindrical target with a diameter of 7 cm and a length of 50 cm can be installed inside a cave under concrete or iron shielding. The beam intensity is measured by a precision ionisation chamber (PIC) and given in PIC counts which are proportional to the number of particles impinging on the secondary copper target (22116 ± 92 particles per PIC count⁽³⁾). Outside the 80 cm thick concrete shielding a mixed radiation field is provided which is sufficiently similar to the one that is to be expected at the LHC. Details about the facility and the reference positions can be found in ref.⁽⁴⁾. During the measurements a set of hydrogen-filled IG5 ionisation chambers was placed at

*Corresponding author: christian.theis@cern.ch

Table 1. Exposure locations used for the IG5 chambers at CERF. The abbreviations, as given in the first column, will be used throughout this paper.

Abbreviation	Description
CS2	Concrete side position 2
CS4	Concrete side position 4
CS-50U	Concrete side position, 50 cm upstream of the target's front face
CT4	Concrete top position 4
CT6/T10	Boundary of concrete top position 6 and 10
CT16	Concrete top position 16

different exposure locations (see Table 1) outside of the concrete shielding and measurements were conducted for beam intensities of 500, 1000, 2000 and 4000 PIC counts.

During the experiment the created charge in the active volume of the ionisation chambers was recorded as a function of the beam intensity and the exposure location. This quantity is converted into counts by a charge digitizer and after the application of the neutron calibration factor (see Table 2), which was previously determined with a ^{238}Pu –Be source, one obtains the corresponding ambient dose equivalent.

During the standard calibration procedure at CERN also a photon calibration factor is determined. But as can be seen from the calculated particle fluence spectra for the CT6/10 position in Figure 1, the dominant field component are neutrons. Consequently, the application of the neutron calibration factor is assumed to be justified. In previous studies⁽⁵⁾ particle fluence spectra were calculated for the respective reference locations using FLUKA. Additionally, energy dependent response functions of the detector were obtained for various particles by FLUKA calculations of the energy deposition within the active volume, and subsequent conversion into charge by the application of the energy required to produce an electron/ion pair (W -factor). Convoluting the particle spectra given at various locations with the response functions yields the total created charge in the active detector volume. Comparing these values to the experimentally obtained detector counts, which correspond directly

to the created charge within the active volume, showed good agreement within 15% (see reference (5)).

The next step was the application of the source-based calibration factor (see Table 2) to the detector readings to calculate ambient dose equivalent. In order to calculate a radiation field dependent calibration factor, a convolution of the calculated particle fluence spectra at the various positions with the corresponding fluence-to-ambient-dose-equivalent conversion factors had to be performed. All results were normalised to a beam intensity of 1000 PIC counts. Table 3 shows the results for the different exposure locations that were obtained from the simulations as well as the average results of the measurements expressed in terms of ambient dose equivalent using the source-based calibration factor. In addition, the respective ratio of the simulation to the measured values are given for the ambient dose equivalent rates, as well as the unbiased, charge-based comparisons of previous studies⁽⁵⁾.

As can be seen from Table 3, the comparison of the simulation and the experiment in terms of ambient dose equivalent shows reasonably good agreement for the top locations. For the side positions of the concrete shielding the simulations give higher results than their measured values. The most significant deviation is visible for the position CS-50U, but this is a geometric effect in the simulations and was explained in detail in ref. (5). The positions CS2 and CT6/T10 should show comparable results because the distance from the ionisation chamber to the target and the amount of shielding in between, was chosen to be identical. However, the simulations as well as the charge measurements (last column in Table 3) showed higher values for the CS2 position. Closer investigation yielded that in comparison to the top locations the calculated particle fluence spectra at the side positions show a significantly higher contribution of low-energetic neutrons (see Figure 2), which can be attributed to back-scattered particles due to an additional shielding wall. The charge-based comparisons take this effect into account correctly and consequently good agreement was observed in the comparison of simulation and measurements. However, this was not the case for the ambient-dose-equivalent-based comparisons as can be seen from the respective ratios in Table 3.

MEASUREMENTS AT THE CERN PS

The PS is the oldest of the accelerators in operation at CERN. Protons are injected at a momentum of 1.4 GeV/c into the PS, where they are accelerated to momenta between 14 GeV/c and 26 GeV/c. The PS has a radius of 100 m. It is installed in a ring-shaped accelerator building, slightly under normal ground level and covered with the excavated soil to form a shielding hill. The shielding of the PS has been

Table 2. Neutron calibration factors including absolute errors for hydrogen-filled IG5 ionisation chambers, which were determined with a standard ^{238}Pu –Be source.

Detector ID	Calibration factor [nSv/count]
H20-1	7.20 ± 0.51
H20-2	6.95 ± 0.50

FIELD CALIBRATION STUDIES

Figure 1. Calculated particle fluence spectra per primary particle at the position CT6/T10 for various particles (lethargy representation)⁽⁵⁾.

designed 50 years ago for a beam intensity which has since been exceeded by approximately a factor of 100. In addition, legal dose limits for personnel and for the public have been lowered, and together with them guideline values for assessing the environmental impact of sources of ionising radiation. The weakest part of the shielding is the so-called 'PS Bridge', 30 m long concrete structure covering the accelerator at the location where it traverses two former experimental halls. The concrete thickness of

the roof over the accelerator measures only between 1.80 and 3.40 m.

For the measurements, two IG5 ionisation chambers were used on the PS Bridge, one filled with hydrogen, and one with argon⁽⁶⁾. Table 4 shows the results of the measurement.

In contrast to the situation at CERF, in the PS the exact location of the beam loss and its intensity are unknown. The loss is approximately situated at the location of maximum activation of the PS vacuum chamber, measured after stop of the accelerator. A loss of protons at an energy $E_{\text{kin}} = 13.1 \text{ GeV}$ ($p = 14 \text{ GeV/c}$) in the centre of a PS magnet unit has been modelled within FLUKA^(1,2). As for CERF, particle fluence spectra of neutrons, photons, protons and pions were scored on top of the PS Bridge at the location of the ionisation chambers. Ambient dose equivalent calculated from fluence spectra and conversion factors⁽⁷⁾ yields a result of $H^*(10) = (2.22 \pm 0.03) \mu\text{Sv}/10^{12}$ protons on the PS Bridge. Additional details of the simulation can be found in ref. (8). The monitor indications are calculated from the response functions from ref. (5). Since information about the intensity of the beam loss is not available, the ratio of the different monitor indications is compared with the measurement. Table 5 shows the calculated values and their ratios.

The observed and calculated ratios of monitor indications from stray radiation at the PS Bridge coincide within their uncertainties, proving that the

Table 3. Ambient dose equivalent for different reference locations at the CERF facility normalised to a beam intensity of 1000 PIC counts. Additionally, the ratios for the calculated versus the measured ambient dose equivalent rates are given, as well as the ratios of the calculated versus the measured charge⁽⁵⁾.

Location	FLUKA $H^*(10)$ [$\mu\text{Sv}/\text{h}/$ 1000 PIC] (%)	IG5 $H^*(10)$ [$\mu\text{Sv}/\text{h}/$ 1000 PIC] (%)	Ratio $H^*(10)$ sim/exp (%)	Ratio charge sim/exp (%)
CS2	121.55 ± 13	95.33 ± 1	1.28 ± 13	1.11 ± 14
CS4	89.68 ± 12	72.93 ± 4	1.23 ± 12	1.23 ± 16
CS-50U	34.54 ± 22	25.96 ± 13	1.33 ± 26	1.38 ± 21
CT4	56.14 ± 9	48.08 ± 5	1.17 ± 10	1.14 ± 15
CT6/T10	94.10 ± 10	94.25 ± 7	1.00 ± 10	1.10 ± 13
CT16	54.08 ± 10	53.25 ± 1	1.02 ± 10	0.96 ± 15

Figure 2. Neutron fluence spectra per primary particle for the different CERF exposure locations that were obtained from FLUKA simulations (lethargy representation).

Table 4. Observed monitor indications on the PS Bridge.

Protons extracted (h^{-1})	Monitor indication IG5-H20	Monitor indication IG5-A20	Ratio of Monitor indications (charge) IG5-H20/IG5-A20
$6.9 \cdot 10^{15}$	$78.2 \pm 5.1 \mu\text{Sv h}^{-1}$	$13.1 \pm 0.1 \mu\text{Sv h}^{-1}$	0.17 ± 0.02
$6.9 \cdot 10^{15}$	$11.3 \pm 1.0 \text{ pC h}^{-1}$	$66.2 \pm 0.5 \text{ pC h}^{-1}$	0.17 ± 0.02

Note: The charge response of the ionisation chambers is calculated from the monitor calibration factor obtained from the ^{238}Pu –Be source. In the last columns, the monitor indication and the generated charge are compared between the two monitors.

Table 5. Calculated charge generated in monitors.

Calculated charge IG5-H20	Calculated charge IG5-A20	Ratio of calculated charges IG5-H20/IG5-A20
$297 \pm 12.5 \text{ pC}/(10^{12}\text{p})$	$1530 \pm 65 \text{ pC}/(10^{12}\text{p})$	0.19 ± 0.01
$2.05 \pm 0.08 \mu\text{Sv}/(10^{12}\text{p})$	$0.31 \pm 0.01 \mu\text{Sv}/(10^{12}\text{p})$	0.19 ± 0.01

Note: The charge generated in the monitors is estimated from calculated response functions and fluence spectra. The collected charge is converted into a dose equivalent indication with the measured calibration factor of the monitor. In the last column, the generated charge is compared between the two monitors.

assumptions on beam loss and the simulations are based on a sufficiently accurate description of the real situation. This allows for comparing the direct Monte Carlo estimation of $H^*(10)$ and the calculated charge response of the monitor, converted into a dose equivalent reading with a measured

calibration factor (Table 6). A hydrogen-filled IG5 ionisation chamber, calibrated with a ^{238}Pu –Be neutron source, indicates ambient dose equivalent in the composite stray radiation field at the PS Bridge with a deviation of only a few percent. The smaller response of the ionisation chamber for neutrons at

Table 6. Comparison of the ‘direct’ Monte-Carlo calculation of ambient dose equivalent and the calculated monitor indication, using the measured calibration coefficient of the IG5-H20 ionisation chamber. The calibration coefficient is suitable for the radiation field at the PS Bridge and the CERN boundary although it has been determined in a different radiation field.

Location	Calculated ambient dose equivalent $H_{MC}^*(10)$	Calculated monitor indication $H_{Mon}^*(10)$	Ratio MC/Monitor
PS Bridge	$(2.22 \pm 0.03) \mu\text{Sv}/(10^{12} \text{ p})$	$(2.05 \pm 0.08) \mu\text{Sv}/(10^{12} \text{ p})$	1.08 ± 0.04

an energy of about 100 MeV (Figure 4) is compensated by the response of the chamber to other particles, mainly photons and protons.

COMPARISON WITH SOURCE CALIBRATION OF THE DETECTOR

Figure 3 illustrates the neutron ambient dose equivalent spectrum obtained for the ^{238}Pu –Be source at CERN⁽⁹⁾ and the CT6/T10 position at CERF. As can be seen from Figure 3 the major contribution to the ambient dose equivalent at CT6/T10 is coming from high-energetic neutrons above 10 MeV. Therefore, it is striking that in general the application of the source-based calibration factor, which does not cover the energies encountered in the radiation field of CERF, still yields reasonably good agreement.

Finally, a calibration factor of the IG5 detector with respect to the ^{238}Pu –Be source was calculated

by the convolution of the neutron fluence spectrum with the respective calculated response function expressed in terms of ambient dose equivalent⁽⁵⁾. The same procedure was performed for different exposure locations at the CERF facility using the respective mixed particle spectra and response functions. Comparing the ^{238}Pu –Be source calibration factors with the calculated field calibration factor for CERF shows good agreement despite differing source spectra (Table 7).

Taking the neutron response of the detector into account one would expect an underestimation as it shows a lower sensitivity to neutrons above 10 MeV than to those below this threshold (see response function in Figure 4). However, in total this is mainly compensated by the contribution of other particles (mainly protons and photons) and thus, fairly good agreement is achieved for this specific field composition. However, one should keep in mind that such a source-based factor is not generally

Figure 3. Neutron ambient dose equivalent spectrum for the ^{238}Pu –Be source measured at CERN (including scattered neutrons which contribute about 30% at the position of calibration) and for the mixed CERF field at the exposure location CT6/T10.

Figure 4. Neutron response of IG5-H20 ionisation chamber (data points, left-hand scale) and fluence spectrum of neutrons on the PS Bridge (histogram, right-hand scale) and. Calibrated with ^{238}Pu -Be, at a neutron energy of approximately 4 MeV, the chamber will underestimate dose equivalent from the high-energy peak in the spectrum. This is partly compensated by the chamber response to other particles, mainly protons and photons.

applicable to every mixed high-energy radiation field, as there is a strong dependence on the fluence spectrum as well as the field composition.

SUMMARY AND CONCLUSIONS

Several measurement campaigns were performed at the CERF facility to test the behaviour of hydrogen-filled IG5 ionisation chambers in a mixed high-energy radiation field. Accompanying FLUKA simulations were performed to obtain the spectral particle fluence, as well as the ambient dose equivalent rates for various exposure locations. Subsequently, the detector readings were converted into dose equivalent values by the application of a ^{238}Pu -Be source-based calibration factor. Even though the neutron energy range covered by the

calibration source does not fully account for the different particle types and the energy range observed at the CERF facility, the results showed reasonably good agreement with the simulations in terms of ambient dose equivalent. This can be explained by the fact that this limitation is recovered by the relatively high sensitivity of the detector in the energy range of the ^{238}Pu -Be source in comparison to neutron energies above 10 MeV, and also because of the contribution of other particles in the mixed field. Still, it should be noted that from this agreement one cannot deduce that a source-based calibration factor is generally applicable to every mixed high-energy radiation field. However, it was shown that in radiation environments outside shielded areas, which exhibit the particle and energy composition comparable to the one encountered at the top locations of CERF, the application of the source-based calibration factor has been appropriate. Nevertheless, improvement can be achieved by calculating field specific calibration functions based on Monte Carlo simulations. Presently, such studies and calculations are under way.

Table 7. Measured and calculated neutron calibration factors for hydrogen-filled IG5 chambers in the field of the ^{238}Pu -Be source at CERN and calculated mixed field calibration factor for the CT6/T10 position at CERF.

	Calibration factor [nSv/pC]
^{238}Pu -Be-H20-1	7.20 ± 0.51
^{238}Pu -Be-H20-2	6.95 ± 0.50
^{238}Pu -Be-FLUKA	6.97 ± 0.34
CERF-FLUKA	6.98 ± 0.36

ACKNOWLEDGEMENTS

The authors would like to express their gratitude to Norbert Aguilar, Pierre Carbonez, Michel Pangallo, Daniel Perrin and Markus Rettig for their support in setting up the experiments.

REFERENCES

1. Fassò, A., Ferrari, A., Ranft, J. and Sala, P. R. *FLUKA: a multi-particle transport code*. CERN Yellow Report (2005), INFN/TC_05/11, SLAC-R-773.
2. Fassò, A., Ferrari, A., Roesler, S., Sala, P. R., Battistoni, G., Cerutti, F., Gadioli, E., Garzelli, M. V., Ballarini, F. and Ottolenghi, A. *et al. The physics models of FLUKA: status and recent developments*. Computing in High Energy and Nuclear Physics 2003 Conference (CHEP2003), La Jolla, CA, USA, March 24–28, 2003 (paper MOMT005) eConf C0303241 (2003), arXiv:hep-ph/0306267.
3. Vincke, H., Mayer, S., Efthymiopoulos, I., Fabich, A., Forkel-Wirth, D., Müller, M. J. and Theis, C. *Accurate PIC calibration by the use of a coincidence of two scintillators*. CERN Technical Note CERN-SC-2004-090-RP-TN (2004).
4. Mitaroff, A. and Silari, M. *The CERN-EU high-energy Reference Field (CERF) facility for dosimetry at commercial flight altitudes and in space*. Radiat. Prot. Dosim. **102**, 7–22 (2002).
5. Theis, C., Forkel-Wirth, D., Perrin, D., Roesler, S. and Vincke, H. *Characterisation of ionisation chambers for a mixed radiation field and investigation of their suitability as radiation monitors for the LHC*. Radiat. Prot. Dosim. **116**, 170–174 (2005).
6. Mayer, S. *Dose equivalent measurements in mixed and time varying radiation fields around high-energy accelerators*. Doctoral thesis, CERN-THESIS-2003-012, (2003).
7. Pelliccioni, M. *Overview of fluence-to-effective dose and fluence-to-ambient dose equivalent conversion coefficients for high energy radiation calculated using the FLUKA code*. Radiat. Prot. Dosim. **88**, 279–297 (2000).
8. Otto, Th. *Beam Loss in the CERN PS and Consequences for Radiation Protection*. CERN-SC-2005-144-RP-TN (2006).
9. Naismith, O. F. and Siebert, B. R. L. *Manual for SPKTBIB*. NPL Report CIRA (EXT) 005 (1996).